
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001158
11:04 08/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:38
STERILIZING 4:51
136.0 C 220kPa 4:51
136.4 C 224kPa 5:21
136.4 C 226kPa 5:51
136.4 C 225kPa 6:21
136.4 C 226kPa 6:51
136.4 C 225kPa 7:21
136.3 C 224kPa 7:51
136.4 C 225kPa 8:21
136.4 C 225kPa 8:21
VENTING 8:21
AIR DRYING 9:16
CYCLE COMPLETE 10:18

Digital Signature #
7949D2FC059631FE
