

-----  
Midmark M11 - v1.0.0  
Total Cycles: 944

Name: RITTER 1  
Sterilizer ID: V2178198  
06/28/2021 06:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.7 F 0.0  
2:00 74.8 F 0.0  
4:00 75.3 F 0.0  
6:00 94.9 F 0.0  
8:00 133.5 F 0.1  
10:00 169.5 F 0.2  
12:00 194.1 F 0.3  
14:00 210.4 F 3.3  
16:00 217.1 F 6.0  
18:00 228.9 F 9.0  
20:00 237.2 F 12.5  
22:00 234.3 F 8.8  
24:00 245.0 F 12.3  
26:00 242.8 F 11.8  
28:00 240.3 F 10.3  
30:00 252.8 F 16.5  
32:00 263.6 F 22.8  
34:00 272.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.6 F 28.6  
2:00 271.3 F 28.0  
3:00 272.7 F 28.7  
4:00 271.4 F 27.9

Min 271.2 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
5:00 222.2 F 0.0  
10:00 263.5 F 0.0  
15:00 240.5 F 0.0  
20:00 203.6 F 0.0  
25:00 199.2 F 0.0  
30:00 191.6 F 0.0

Drying Complete

FILLING: 0:45  
HEATING: 34:15  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:12:15  
CYCLE COMPLETE

06/28/2021 08:15 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01