

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4028

Name: RITTER2  
Sterilizer ID: V2313228  
10/15/2024 11:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 97.9 F 0.0  
2:00 99.4 F 0.0  
4:00 119.7 F 0.0  
6:00 156.2 F 0.6  
8:00 170.0 F 3.2  
10:00 191.9 F 6.6  
12:00 202.0 F 9.9  
14:00 214.9 F 13.5  
16:00 222.3 F 7.0  
18:00 231.5 F 11.3  
20:00 230.9 F 7.9  
22:00 243.1 F 13.2  
24:00 242.0 F 11.8  
26:00 253.3 F 17.5  
28:00 263.5 F 23.5  
30:00 272.8 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 272.2 F 29.6  
2:00 272.4 F 29.3  
3:00 271.5 F 29.1  
4:00 272.9 F 29.8

Min 271.5 F 28.9  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
5:00 188.5 F 0.0  
10:00 186.8 F 0.0  
15:00 184.5 F 0.0  
20:00 183.9 F 0.0  
25:00 190.5 F 0.0  
30:00 190.0 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 30:14  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:08:52  
CYCLE COMPLETE

10/15/2024 12:17 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_