
Midmark M11 - v1.0.5
Total Cycles: 1367

Name: RITTER2
Sterilizer ID: V2313228
08/04/2022 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.5 F 0.0
2:00 125.7 F 0.0
4:00 129.8 F 0.1
6:00 168.1 F 1.4
8:00 187.5 F 5.7
10:00 210.6 F 10.2
12:00 222.3 F 6.6
14:00 231.5 F 10.4
16:00 233.7 F 8.3
18:00 243.6 F 13.4
20:00 243.0 F 12.3
22:00 255.2 F 18.7
24:00 266.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 30.0
2:00 272.2 F 29.3
3:00 271.3 F 29.1
4:00 272.4 F 30.0

Min 271.2 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 211.0 F 0.0
10:00 225.2 F 0.0
15:00 227.0 F 0.0
20:00 222.2 F 0.0
25:00 220.1 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:44
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

08/04/2022 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____