
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 004759
15:18 22/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:26
STERILIZING 3:22
136.5 C 222kPa 3:22
136.3 C 223kPa 3:53
136.3 C 224kPa 4:23
136.4 C 225kPa 4:53
136.4 C 223kPa 5:23
136.4 C 224kPa 5:53
136.4 C 224kPa 6:23
136.4 C 225kPa 6:53
136.4 C 225kPa 6:53
VENTING 6:53
AIR DRYING 7:43
CYCLE COMPLETE 8:08

Digital Signature #
1D5264E6594B2D75
