
Midmark M11 - v1.0.0
Total Cycles: 1256

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2021 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.6 F 0.0
2:00 105.3 F 0.0
4:00 126.3 F 0.1
6:00 160.6 F 0.1
8:00 185.2 F 2.1
10:00 204.3 F 5.6
12:00 216.5 F 9.1
14:00 227.7 F 13.0
16:00 229.9 F 8.1
18:00 241.0 F 12.8
20:00 238.9 F 10.3
22:00 236.4 F 8.6
24:00 249.2 F 14.8
26:00 260.6 F 21.1
28:00 271.0 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.6 F 28.2
3:00 272.0 F 28.8
4:00 272.4 F 28.7

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 213.5 F 0.0
10:00 215.6 F 0.0
15:00 211.2 F 0.0
20:00 206.7 F 0.0
25:00 200.5 F 0.0
30:00 197.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:46
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:06:56
CYCLE COMPLETE

10/12/2021 02:21 PM
ACCEPTED: X
REJECTED:
Operator: U01