
Midmark M11 - v1.0.0
Total Cycles: 1698

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2022 08:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 75.3 F 0.0
4:00 106.3 F 0.2
6:00 138.5 F 1.5
8:00 154.2 F 2.4
10:00 175.5 F 4.6
12:00 191.0 F 7.6
14:00 205.5 F 11.2
16:00 214.4 F 4.5
18:00 225.3 F 7.5
20:00 235.5 F 11.5
22:00 232.9 F 7.8
24:00 245.5 F 13.5
26:00 243.4 F 11.8
28:00 255.6 F 18.0
30:00 266.2 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.4
2:00 272.3 F 28.7
3:00 271.6 F 28.0
4:00 272.7 F 28.8

Min 271.4 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.9 F 0.0
5:00 196.4 F 0.0
10:00 206.9 F 0.0
15:00 197.2 F 0.0
20:00 196.6 F 0.0
25:00 196.0 F 0.0
30:00 196.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:43
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:09:35
CYCLE COMPLETE

03/09/2022 09:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____