

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3089

Name: RITTER2  
Sterilizer ID: V2313228  
11/06/2023 02:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.8 F 0.0  
2:00 137.0 F 1.5  
4:00 163.0 F 4.7  
6:00 185.9 F 9.2  
8:00 207.5 F 7.2  
10:00 220.6 F 8.2  
12:00 234.0 F 13.3  
14:00 236.9 F 10.6  
16:00 235.7 F 9.1  
18:00 250.6 F 16.1  
20:00 262.8 F 23.2  
22:00 272.2 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.5 F 29.7  
2:00 271.5 F 29.1  
3:00 272.5 F 29.8  
4:00 271.8 F 29.1

Min 271.3 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 201.3 F 0.0  
10:00 201.9 F 0.0  
15:00 199.4 F 0.0  
20:00 196.7 F 0.0  
25:00 194.9 F 0.0  
30:00 194.4 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 22:11  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:00:36  
CYCLE COMPLETE

11/06/2023 03:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_