
Midmark M11 - v1.0.0
Total Cycles: 3751

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2023 01:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.6 F 0.0
2:00 163.4 F 0.1
4:00 183.4 F 2.9
6:00 202.5 F 6.3
8:00 217.2 F 10.7
10:00 227.4 F 6.2
12:00 237.0 F 10.8
14:00 234.1 F 7.6
16:00 245.4 F 12.4
18:00 245.4 F 12.7
20:00 257.2 F 18.9
22:00 267.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.6 F 29.0
2:00 272.0 F 28.7
3:00 272.1 F 28.4
4:00 272.7 F 28.9

Min 270.9 F 27.9
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.7 F 0.0
5:00 238.0 F 0.0
10:00 232.0 F 0.0
15:00 231.0 F 0.0
20:00 226.4 F 0.0
25:00 225.2 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:15
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:48
CYCLE COMPLETE

09/11/2023 02:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____