
Midmark M11 - v1.0.5
Total Cycles: 2444

Name: RITTER2
Sterilizer ID: V2313228
05/31/2023 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.0 F 0.0
2:00 199.6 F 4.2
4:00 205.9 F 9.6
6:00 222.9 F 6.9
8:00 233.6 F 11.3
10:00 235.3 F 9.2
12:00 237.6 F 9.4
14:00 249.7 F 15.6
16:00 262.3 F 22.8
18:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.9
2:00 272.2 F 29.2
3:00 271.4 F 28.9
4:00 272.3 F 29.9

Min 271.3 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 221.0 F 0.0
10:00 235.1 F 0.0
15:00 232.1 F 0.0
20:00 222.2 F 0.0
25:00 220.0 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

05/31/2023 12:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____