
Midmark M11 - v1.0.5
Total Cycles: 3314

Name: RITTER2
Sterilizer ID: V2313228
01/24/2024 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.1 F 0.0
2:00 153.4 F 1.2
4:00 166.2 F 4.1
6:00 191.2 F 8.3
8:00 209.7 F 13.3
10:00 224.3 F 7.4
12:00 237.4 F 12.3
14:00 235.3 F 9.3
16:00 239.4 F 9.7
18:00 246.9 F 13.8
20:00 258.5 F 20.2
22:00 269.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.4
2:00 271.7 F 28.6
3:00 272.5 F 29.6
4:00 272.4 F 29.0

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 225.7 F 0.0
10:00 202.6 F 0.0
15:00 198.6 F 0.0
20:00 195.7 F 0.0
25:00 193.2 F 0.0
30:00 191.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 23:09
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

01/24/2024 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____