
Midmark M11 - v1.0.0
Total Cycles: 1585

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2022 04:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 151.8 F 2.7
4:00 170.6 F 6.7
6:00 209.0 F 11.5
8:00 215.5 F 4.9
10:00 229.5 F 9.9
12:00 234.7 F 8.4
14:00 242.3 F 12.3
16:00 240.7 F 10.7
18:00 254.0 F 17.4
20:00 265.5 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.4 F 29.0
2:00 272.0 F 28.5
3:00 272.2 F 29.0
4:00 272.3 F 28.8

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 231.6 F 0.0
10:00 228.0 F 0.0
15:00 226.9 F 0.0
20:00 226.8 F 0.0
25:00 228.6 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:36
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

02/01/2022 05:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____