
Midmark M11 - v1.0.0
Total Cycles: 1356

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2021 03:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.4 F 0.0
2:00 211.9 F 3.8
4:00 217.0 F 8.8
6:00 230.9 F 9.4
8:00 235.4 F 10.1
10:00 237.7 F 8.1
12:00 244.5 F 12.4
14:00 245.0 F 12.6
16:00 257.6 F 19.3
18:00 268.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.2 F 29.0
2:00 271.8 F 28.3
3:00 271.8 F 28.8
4:00 272.2 F 28.7

Min 271.3 F 28.1
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.0
5:00 230.9 F 0.0
10:00 238.4 F 0.0
15:00 235.7 F 0.0
20:00 226.9 F 0.0
25:00 227.8 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 3:08
HEATING: 19:04
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

11/10/2021 04:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____