

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000958  
15:12 02/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:56  
PRESSURIZING 4:57  
STERILIZING 6:12  
136.0 C 229kPa 6:12  
136.4 C 233kPa 6:43  
136.4 C 232kPa 7:13  
136.4 C 232kPa 7:43  
136.2 C 230kPa 8:13  
136.3 C 231kPa 8:43  
136.4 C 233kPa 9:13  
136.4 C 232kPa 9:43  
136.4 C 232kPa 9:43  
VENTING 9:43  
AIR DRYING 10:41  
CYCLE COMPLETE 25:41  
-----

Digital Signature #  
7219740D3FC2BAF8  
-----