
Midmark M11 - v1.0.5
Total Cycles: 1388

Name: RITTER2
Sterilizer ID: V2313228
08/15/2022 07:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 87.7 F 0.0
4:00 81.9 F 0.0
6:00 109.2 F 0.0
8:00 132.5 F 1.1
10:00 161.8 F 4.1
12:00 183.5 F 7.6
14:00 205.7 F 11.1
16:00 218.2 F 5.8
18:00 230.8 F 8.7
20:00 238.4 F 12.8
22:00 236.0 F 9.2
24:00 240.2 F 10.8
26:00 244.7 F 12.7
28:00 255.4 F 18.4
30:00 265.3 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.2 F 29.5
2:00 271.8 F 29.1
3:00 272.7 F 29.3
4:00 272.5 F 29.8

Min 270.8 F 28.5
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 220.3 F 0.0
10:00 232.2 F 0.0
15:00 243.4 F 0.0
20:00 231.5 F 0.0
25:00 229.3 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:53
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:52
CYCLE COMPLETE

08/15/2022 08:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____