
Midmark M11 - v1.0.0
Total Cycles: 1269

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2021 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.8 F 0.0
2:00 190.4 F 4.1
4:00 214.3 F 10.4
6:00 226.7 F 6.2
8:00 241.1 F 12.5
10:00 240.7 F 10.7
12:00 238.7 F 9.7
14:00 254.1 F 17.2
16:00 266.9 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.0 F 28.2
2:00 272.4 F 28.7
3:00 271.3 F 28.2
4:00 272.2 F 28.4

Min 271.2 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 234.0 F 0.0
10:00 221.0 F 0.0
15:00 220.9 F 0.0
20:00 222.9 F 0.0
25:00 211.9 F 0.0
30:00 198.9 F 0.0
Drying Complete

FILLING: 3:05
HEATING: 17:20
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:32
CYCLE COMPLETE

10/14/2021 02:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____