
Midmark M11 - v1.0.0
Total Cycles: 2893

Name: RITTER 1
Sterilizer ID: V2178198
02/13/2023 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.2 F 0.0
2:00 194.6 F 4.3
4:00 197.1 F 9.2
6:00 216.0 F 8.0
8:00 228.8 F 9.6
10:00 233.0 F 8.7
12:00 243.3 F 13.0
14:00 242.4 F 11.8
16:00 255.5 F 18.6
18:00 267.0 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.4
2:00 271.7 F 28.7
3:00 271.7 F 29.1
4:00 272.4 F 29.4

Min 271.2 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 240.1 F 0.0
10:00 231.2 F 0.0
15:00 228.7 F 0.0
20:00 230.4 F 0.0
25:00 229.4 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

02/13/2023 04:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____