
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004051
10:17 19/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:56
STERILIZING 4:13
136.4 C 223kPa 4:13
136.4 C 223kPa 4:43
136.4 C 223kPa 5:13
136.4 C 223kPa 5:43
136.2 C 222kPa 6:13
136.3 C 223kPa 6:43
136.3 C 221kPa 7:13
136.4 C 223kPa 7:43
136.4 C 223kPa 7:43
VENTING 7:43
AIR DRYING 8:41
CYCLE COMPLETE 9:04

Digital Signature #
905620F6D8E23726
