
Midmark M11 - v1.0.5
Total Cycles: 2409

Name: RITTER2
Sterilizer ID: V2313228
05/24/2023 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.7 F 0.0
2:00 189.6 F 3.8
4:00 207.0 F 9.3
6:00 222.5 F 7.5
8:00 235.8 F 11.6
10:00 236.8 F 9.9
12:00 235.9 F 9.0
14:00 251.9 F 16.6
16:00 264.6 F 24.1
18:00 272.5 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.5 F 29.6
2:00 272.7 F 30.1
3:00 271.9 F 29.2
4:00 273.0 F 30.0

Min 271.7 F 29.2
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 225.7 F 0.0
10:00 244.6 F 0.0
15:00 230.1 F 0.0
20:00 224.6 F 0.0
25:00 221.2 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

05/24/2023 02:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____