

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001407  
13:17 09/08/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:57  
PRESSURIZING 5:25  
STERILIZING 6:42  
136.0 C 220kPa 6:42  
136.3 C 223kPa 7:12  
136.5 C 224kPa 7:42  
136.0 C 220kPa 8:12  
136.4 C 222kPa 8:42  
136.3 C 223kPa 9:12  
136.2 C 222kPa 9:42  
136.3 C 222kPa 10:12  
VENTING 10:12  
AIR DRYING 11:20  
CYCLE COMPLETE 11:25  
-----

Digital Signature #  
0F574DCAA4EDB6D5  
-----