

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1926

Name: RITTER2  
Sterilizer ID: V2313228  
01/12/2023 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.5 F 0.0  
2:00 129.9 F 1.4  
4:00 144.6 F 4.0  
6:00 167.5 F 7.8  
8:00 193.6 F 12.9  
10:00 214.3 F 6.5  
12:00 227.3 F 11.6  
14:00 231.2 F 8.4  
16:00 240.5 F 11.2  
18:00 244.6 F 13.3  
20:00 257.3 F 20.1  
22:00 268.8 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 273.0 F 30.1  
2:00 272.1 F 29.5  
3:00 271.4 F 29.4  
4:00 272.5 F 30.2

Min 271.3 F 29.1  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
5:00 225.9 F 0.0  
10:00 231.6 F 0.0  
15:00 228.0 F 0.0  
20:00 225.6 F 0.0  
25:00 227.9 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 23:10  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:01:13  
CYCLE COMPLETE

01/12/2023 11:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_