
Midmark M11 - v1.0.5
Total Cycles: 329

Name: RITTER2
Sterilizer ID: V2313228
09/02/2021 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.1 F 0.0
2:00 155.5 F 2.1
4:00 167.0 F 5.9
6:00 191.8 F 10.9
8:00 211.7 F 4.7
10:00 224.3 F 10.1
12:00 233.5 F 8.1
14:00 241.7 F 12.8
16:00 241.5 F 11.6
18:00 253.6 F 18.1
20:00 265.0 F 24.9
22:00 273.0 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.7
1:00 272.4 F 30.1
2:00 272.3 F 29.9
3:00 272.0 F 29.8
4:00 271.9 F 29.7

Min 271.8 F 29.6
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.3 F 0.0
5:00 233.3 F 0.0
10:00 219.9 F 0.0
15:00 209.3 F 0.0
20:00 184.8 F 0.0
25:00 178.9 F 0.0
30:00 180.6 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

09/02/2021 09:15 AM
ACCEPTED: X
REJECTED:
Operator: U02