

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4496

Name: RITTER 1  
Sterilizer ID: V2178198  
04/26/2024 11:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.3 F 0.0  
2:00 122.4 F 1.3  
4:00 138.0 F 3.7  
6:00 171.2 F 6.8  
8:00 185.2 F 10.4  
10:00 217.6 F 5.8  
12:00 215.7 F 7.5  
14:00 227.3 F 11.7  
16:00 230.1 F 7.2  
18:00 242.1 F 12.4  
20:00 239.2 F 10.1  
22:00 251.3 F 15.9  
24:00 261.8 F 21.9  
26:00 271.4 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 271.7 F 28.9  
2:00 271.7 F 29.1  
3:00 271.6 F 29.2  
4:00 271.6 F 29.2

Min 271.5 F 28.7  
Max 272.5 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
5:00 207.9 F 0.0  
10:00 234.2 F 0.0  
15:00 217.0 F 0.0  
20:00 214.8 F 0.0  
25:00 215.6 F 0.0  
30:00 219.9 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 27:21  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:06:11  
CYCLE COMPLETE

04/26/2024 12:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_