

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4020

Name: RITTER2  
Sterilizer ID: V2313228  
10/10/2024 02:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.0 F 0.0  
2:00 140.1 F 1.9  
4:00 158.8 F 5.0  
6:00 184.7 F 8.8  
8:00 200.9 F 12.9  
10:00 216.2 F 5.6  
12:00 226.6 F 9.6  
14:00 233.3 F 10.1  
16:00 239.2 F 11.5  
18:00 238.2 F 10.0  
20:00 250.6 F 16.1  
22:00 261.4 F 22.3  
24:00 271.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.1  
1:00 272.0 F 29.7  
2:00 272.4 F 29.5  
3:00 271.3 F 29.0  
4:00 272.1 F 29.9

Min 271.3 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.7 F 0.0  
5:00 198.8 F 0.0  
10:00 198.5 F 0.0  
15:00 196.9 F 0.0  
20:00 195.7 F 0.0  
25:00 193.3 F 0.0  
30:00 190.3 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 24:33  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:03:32  
CYCLE COMPLETE

10/10/2024 03:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_