
Midmark M11 - v1.0.0
Total Cycles: 3933

Name: RITTER 1
Sterilizer ID: V2178198
10/23/2023 08:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 72.7 F 0.0
4:00 81.2 F 0.0
6:00 111.7 F 0.0
8:00 149.8 F 0.3
10:00 176.7 F 3.5
12:00 193.5 F 6.8
14:00 207.7 F 10.1
16:00 217.6 F 11.2
18:00 222.5 F 6.9
20:00 233.5 F 11.0
22:00 230.9 F 7.1
24:00 243.0 F 12.5
26:00 240.7 F 10.7
28:00 253.3 F 17.0
30:00 264.3 F 23.4
32:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.8
2:00 271.5 F 28.4
3:00 272.6 F 28.8
4:00 271.5 F 28.2

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 199.6 F 0.0
10:00 214.9 F 0.0
15:00 225.6 F 0.0
20:00 229.9 F 0.0
25:00 227.8 F 0.0
30:00 221.7 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 32:04
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:10:21
CYCLE COMPLETE

10/23/2023 09:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____