
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001569
14:20 07/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
CYCLE ABORTED 3:12
WATER LEVEL LOW
NOT STERILE
END TIME 3:12

Digital Signature #
D871CED92BD1E36D
