
Midmark M11 - v1.0.5
Total Cycles: 2879

Name: RITTER2
Sterilizer ID: V2313228
09/13/2023 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.9 F 0.0
2:00 124.2 F 0.0
4:00 154.4 F 0.7
6:00 178.1 F 3.9
8:00 195.3 F 7.4
10:00 210.6 F 11.3
12:00 220.2 F 5.0
14:00 232.0 F 9.4
16:00 238.1 F 11.3
18:00 241.0 F 11.5
20:00 239.1 F 10.2
22:00 251.3 F 16.3
24:00 262.1 F 22.5
26:00 272.1 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 271.9 F 29.3
2:00 272.4 F 29.3
3:00 271.4 F 28.7
4:00 271.8 F 29.3

Min 271.2 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 230.4 F 0.0
10:00 188.7 F 0.0
15:00 186.7 F 0.0
20:00 184.2 F 0.0
25:00 182.3 F 0.0
30:00 182.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:28
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:11
CYCLE COMPLETE

09/13/2023 04:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____