
Midmark M11 - v1.0.0
Total Cycles: 1434

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2021 01:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.1 F 0.0
2:00 184.9 F 2.4
4:00 186.8 F 6.1
6:00 209.0 F 11.1
8:00 221.0 F 5.3
10:00 230.0 F 10.4
12:00 233.6 F 7.5
14:00 243.6 F 13.1
16:00 242.9 F 11.7
18:00 255.9 F 18.4
20:00 267.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 28.9
2:00 271.5 F 28.5
3:00 272.1 F 28.8
4:00 271.3 F 28.4

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 206.2 F 0.0
10:00 209.7 F 0.0
15:00 210.4 F 0.0
20:00 209.4 F 0.0
25:00 208.6 F 0.0
30:00 208.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

12/08/2021 02:43 PM
ACCEPTED: X
REJECTED:
Operator: U02