
Midmark M11 - v1.0.5
Total Cycles: 2680

Name: RITTER2
Sterilizer ID: V2313228
07/21/2023 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.6 F 0.0
2:00 110.6 F 0.0
4:00 122.6 F 0.0
6:00 160.9 F 1.1
8:00 182.7 F 4.1
10:00 191.1 F 7.6
12:00 205.3 F 11.3
14:00 220.3 F 5.6
16:00 228.6 F 9.6
18:00 234.4 F 10.6
20:00 239.7 F 11.5
22:00 237.4 F 9.8
24:00 249.6 F 15.7
26:00 260.2 F 21.6
28:00 269.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.2 F 29.8
2:00 272.6 F 29.6
3:00 271.6 F 29.2
4:00 271.0 F 29.0

Min 271.0 F 28.9
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 206.2 F 0.0
10:00 207.6 F 0.0
15:00 208.5 F 0.0
20:00 207.3 F 0.0
25:00 203.4 F 0.0
30:00 203.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:07
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:07:29
CYCLE COMPLETE

07/21/2023 02:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____