
Midmark M11 - v1.0.5
Total Cycles: 3687

Name: RITTER2
Sterilizer ID: V2313228
06/12/2024 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 79.7 F 0.0
4:00 75.1 F 0.5
6:00 105.3 F 2.3
8:00 140.5 F 5.2
10:00 169.2 F 8.1
12:00 202.8 F 11.6
14:00 209.6 F 3.8
16:00 219.9 F 7.2
18:00 231.6 F 10.8
20:00 228.9 F 7.4
22:00 237.6 F 10.7
24:00 235.0 F 8.2
26:00 246.6 F 14.0
28:00 257.7 F 19.9
30:00 267.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.2
2:00 272.4 F 29.7
3:00 271.6 F 28.8
4:00 272.8 F 29.5

Min 271.4 F 28.8
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 223.6 F 0.0
10:00 200.5 F 0.0
15:00 188.4 F 0.0
20:00 189.5 F 0.0
25:00 185.4 F 0.0
30:00 184.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:20
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

06/12/2024 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____