
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001608
10:05 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:27
PRESSURIZING 1:33
STERILIZING 2:42
136.0 C 220kPa 2:42
136.4 C 224kPa 3:12
136.3 C 223kPa 3:42
136.2 C 222kPa 4:12
136.4 C 224kPa 4:42
136.4 C 224kPa 5:12
136.1 C 222kPa 5:42
136.3 C 223kPa 6:12
136.3 C 223kPa 6:12
VENTING 6:12
AIR DRYING 7:06
CYCLE COMPLETE 7:14

Digital Signature #
7E7F5DFB787240D6
