

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4515

Name: RITTER 1  
Sterilizer ID: V2178198  
05/02/2024 09:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
2:00 186.0 F 2.6  
4:00 189.0 F 6.2  
6:00 202.8 F 10.2  
8:00 213.5 F 6.2  
10:00 225.5 F 7.9  
12:00 236.5 F 12.3  
14:00 235.2 F 9.1  
16:00 240.3 F 10.5  
18:00 245.8 F 13.3  
20:00 257.1 F 19.3  
22:00 267.2 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.7 F 29.1  
2:00 271.6 F 28.3  
3:00 272.3 F 29.2  
4:00 272.2 F 28.8

Min 271.4 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 194.6 F 0.0  
10:00 192.0 F 0.0  
15:00 190.6 F 0.0  
20:00 190.8 F 0.0  
25:00 195.5 F 0.0  
30:00 198.2 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 23:30  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:02:56  
CYCLE COMPLETE

05/02/2024 10:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_