
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002126
11:55 04/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:57
STERILIZING 3:05
136.4 C 226kPa 3:05
136.4 C 226kPa 3:35
136.2 C 224kPa 4:05
136.0 C 222kPa 4:35
136.3 C 225kPa 5:05
136.2 C 225kPa 5:35
136.2 C 225kPa 6:05
136.0 C 224kPa 6:35
136.0 C 224kPa 6:35
VENTING 6:36
AIR DRYING 7:25
CYCLE COMPLETE 37:25

Digital Signature #
CBABF561E7C3E3AE
