

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1506

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2022 02:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 214.5 F 0.0  
2:00 199.2 F 4.1  
4:00 217.3 F 9.2  
6:00 226.7 F 6.8  
8:00 235.1 F 10.2  
10:00 234.8 F 7.9  
12:00 246.1 F 13.1  
14:00 244.6 F 12.3  
16:00 257.0 F 18.9  
18:00 268.1 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 271.6 F 28.2  
2:00 272.1 F 28.7  
3:00 271.8 F 28.7  
4:00 271.4 F 28.1  
  
Min 271.2 F 28.0  
Max 272.4 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
5:00 239.8 F 0.0  
10:00 237.0 F 0.0  
15:00 227.3 F 0.0  
20:00 220.5 F 0.0  
25:00 222.3 F 0.0  
30:00 225.3 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 19:13  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:24  
CYCLE COMPLETE

01/10/2022 03:32 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_