
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001751
9:30 24/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:56
STERILIZING 4:09
136.0 C 220kPa 4:09
135.8 C 219kPa 4:39
136.4 C 222kPa 5:09
136.4 C 224kPa 5:39
136.3 C 223kPa 6:09
136.0 C 222kPa 6:39
136.4 C 224kPa 7:09
136.4 C 223kPa 7:39
136.4 C 223kPa 7:39
VENTING 7:39
AIR DRYING 8:27
CYCLE COMPLETE 26:46

Digital Signature #
85A3F39DC613F122
