
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006751
8:44 18/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:31
STERILIZING 4:31
136.0 C 220kPa 4:31
135.9 C 220kPa 5:01
136.2 C 221kPa 5:31
136.4 C 222kPa 6:01
136.4 C 222kPa 6:31
136.2 C 221kPa 7:01
136.4 C 221kPa 7:31
136.2 C 222kPa 8:01
136.2 C 222kPa 8:01
VENTING 8:01
AIR DRYING 8:57
CYCLE COMPLETE 10:27

Digital Signature #
15ABC38B5F1648EF
