
Midmark M11 - v1.0.0
Total Cycles: 4121

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2023 05:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.9 F 0.0
2:00 164.8 F 3.1
4:00 191.0 F 6.9
6:00 204.6 F 11.5
8:00 222.8 F 5.1
10:00 231.5 F 9.5
12:00 236.0 F 9.6
14:00 241.5 F 11.5
16:00 239.2 F 9.8
18:00 252.1 F 16.2
20:00 263.3 F 22.6
22:00 272.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 271.7 F 28.5
2:00 273.0 F 29.0
3:00 271.6 F 28.3
4:00 272.9 F 29.1

Min 270.9 F 27.9
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 221.1 F 0.0
10:00 226.3 F 0.0
15:00 207.0 F 0.0
20:00 198.7 F 0.0
25:00 197.2 F 0.0
30:00 205.1 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 22:09
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

12/14/2023 06:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____