
Midmark M11 - v1.0.5
Total Cycles: 629

Name: RITTER2
Sterilizer ID: V2313228
12/14/2021 07:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.0 F 0.0
2:00 74.5 F 0.0
4:00 82.8 F 0.1
6:00 112.4 F 0.2
8:00 157.8 F 1.8
10:00 165.2 F 5.2
12:00 186.4 F 8.8
14:00 204.6 F 12.6
16:00 218.6 F 5.9
18:00 228.0 F 10.0
20:00 235.8 F 9.5
22:00 240.5 F 11.7
24:00 239.4 F 10.2
26:00 251.8 F 16.4
28:00 262.5 F 22.7
30:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 271.7 F 29.2
2:00 272.1 F 29.0
3:00 271.2 F 28.8
4:00 272.4 F 29.4

Min 271.1 F 28.5
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 215.2 F 0.0
10:00 218.4 F 0.0
15:00 203.7 F 0.0
20:00 210.9 F 0.0
25:00 220.0 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 30:17
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:27
CYCLE COMPLETE

12/14/2021 08:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____