
Midmark M11 - v1.0.5
Total Cycles: 2834

Name: RITTER2
Sterilizer ID: V2313228
09/06/2023 08:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 70.7 F 0.0
4:00 78.6 F 0.0
6:00 101.4 F 0.0
8:00 141.0 F 0.6
10:00 164.2 F 3.6
12:00 184.7 F 7.0
14:00 206.8 F 10.6
16:00 217.6 F 6.8
18:00 226.0 F 8.3
20:00 235.1 F 12.6
22:00 235.5 F 9.6
24:00 238.4 F 9.7
26:00 246.6 F 14.0
28:00 258.0 F 20.2
30:00 268.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.5
2:00 271.5 F 29.0
3:00 272.5 F 29.6
4:00 271.6 F 28.8

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 196.9 F 0.0
10:00 202.1 F 0.0
15:00 202.5 F 0.0
20:00 201.6 F 0.0
25:00 202.2 F 0.0
30:00 202.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:16
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:20
CYCLE COMPLETE

09/06/2023 09:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____