
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004344
11:39 13/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:09
STERILIZING 4:10
136.4 C 224kPa 4:10
136.2 C 224kPa 4:41
136.4 C 221kPa 5:11
136.4 C 224kPa 5:41
136.3 C 223kPa 6:11
136.2 C 222kPa 6:41
136.1 C 223kPa 7:11
136.3 C 225kPa 7:41
136.3 C 225kPa 7:41
VENTING 7:41
AIR DRYING 8:33
CYCLE COMPLETE 8:42

Digital Signature #
0C45209235E2D24F
