
Midmark M11 - v1.0.5
Total Cycles: 4483

Name: RITTER2
Sterilizer ID: V2313228
04/03/2025 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.3 F 0.0
2:00 167.3 F 2.3
4:00 174.4 F 5.8
6:00 195.5 F 10.2
8:00 211.5 F 6.8
10:00 219.8 F 8.6
12:00 229.5 F 13.0
14:00 235.3 F 10.0
16:00 237.2 F 9.3
18:00 246.6 F 14.2
20:00 257.3 F 20.2
22:00 267.7 F 26.7
24:00 273.1 F 30.7

STERILIZING
mm:ss Degrees PSI
0:00 273.0 F 30.2
1:00 272.4 F 30.4
2:00 272.1 F 30.7
3:00 272.1 F 30.5
4:00 272.1 F 30.1

Min 271.9 F 30.1
Max 273.0 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 211.0 F 0.0
10:00 205.8 F 0.0
15:00 183.4 F 0.0
20:00 186.6 F 0.0
25:00 183.4 F 0.0
30:00 184.0 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:03:20
CYCLE COMPLETE

04/03/2025 01:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____