
Midmark M11 - v1.0.0
Total Cycles: 5133

Name: RITTER 1
Sterilizer ID: V2178198
11/26/2024 03:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.1 F 0.0
2:00 87.0 F 0.0
4:00 110.3 F 0.0
6:00 151.4 F 0.2
8:00 181.4 F 0.5
10:00 201.4 F 0.9
12:00 214.2 F 3.8
14:00 224.2 F 7.0
16:00 234.1 F 11.0
18:00 230.9 F 7.3
20:00 244.3 F 13.0
22:00 242.3 F 11.5
24:00 240.4 F 10.5
26:00 253.4 F 16.9
28:00 264.6 F 23.6
30:00 271.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.2 F 28.7
2:00 272.8 F 28.9
3:00 272.4 F 28.6
4:00 271.9 F 28.3

Min 271.6 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.9 F 0.0
5:00 208.1 F 0.0
10:00 205.5 F 0.0
15:00 204.2 F 0.0
20:00 201.7 F 0.0
25:00 200.6 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:06
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

11/26/2024 04:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____