

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 001445  
13:17 07/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:38  
STERILIZING 3:47  
136.4 C 225kPa 3:47  
136.3 C 225kPa 4:17  
136.4 C 226kPa 4:47  
135.9 C 220kPa 5:17  
136.2 C 224kPa 5:47  
136.2 C 224kPa 6:17  
136.4 C 226kPa 6:47  
136.4 C 225kPa 7:17  
136.4 C 225kPa 7:17  
VENTING 7:18  
AIR DRYING 8:07  
CYCLE COMPLETE 23:07  
-----

Digital Signature #  
1C673F0929E62FA7  
-----