

-----  
Midmark M11 - v1.0.0  
Total Cycles: 677

Name: MIDMARK  
Sterilizer ID: V2178198  
02/18/2021 01:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.8 F 0.0  
2:00 133.4 F 0.0  
4:00 126.4 F 0.1  
6:00 159.8 F 1.9  
8:00 191.4 F 5.8  
10:00 210.4 F 10.3  
12:00 229.2 F 6.2  
14:00 229.5 F 9.8  
16:00 235.2 F 8.4  
18:00 243.8 F 12.8  
20:00 242.6 F 11.2  
22:00 256.1 F 18.1  
24:00 267.9 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.8  
2:00 271.9 F 28.7  
3:00 272.5 F 28.7  
4:00 271.8 F 28.7

Min 271.5 F 28.2  
Max 273.1 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 170.3 F 0.0  
5:00 238.8 F 0.0  
10:00 231.9 F 0.0  
15:00 233.4 F 0.0  
20:00 226.6 F 0.0  
25:00 227.3 F 0.0  
30:00 220.6 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 25:58  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 01:04:00  
CYCLE COMPLETE

02/18/2021 02:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_