

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3437

Name: RITTER 1  
Sterilizer ID: V2178198  
06/23/2023 01:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
2:00 182.5 F 3.3  
4:00 194.6 F 7.3  
6:00 217.3 F 13.0  
8:00 224.9 F 7.2  
10:00 237.9 F 12.5  
12:00 237.3 F 9.8  
14:00 236.3 F 8.5  
16:00 249.9 F 15.2  
18:00 262.0 F 22.1  
20:00 272.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.9  
2:00 271.5 F 28.2  
3:00 272.7 F 29.0  
4:00 271.8 F 28.4  
  
Min 271.4 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 218.5 F 0.0  
10:00 235.5 F 0.0  
15:00 232.8 F 0.0  
20:00 240.4 F 0.0  
25:00 223.0 F 0.0  
30:00 207.2 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:45  
CYCLE COMPLETE

06/23/2023 02:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_