
Midmark M11 - v1.0.5
Total Cycles: 178

Name: RITTER2
Sterilizer ID: V2313228
07/07/2021 04:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 173.8 F 2.8
4:00 173.8 F 6.4
6:00 191.8 F 10.4
8:00 204.3 F 5.9
10:00 223.0 F 8.4
12:00 238.1 F 13.1
14:00 240.0 F 11.0
16:00 237.0 F 9.8
18:00 250.7 F 15.9
20:00 261.4 F 22.0
22:00 271.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 272.6 F 29.7
2:00 271.7 F 29.3
3:00 272.4 F 29.3
4:00 272.6 F 29.7

Min 270.9 F 28.6
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 192.0 F 0.0
Dry Cycle Aborted

FILLING: 1:37
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:34
DRYING: 7:46
TOTAL CYCLE: 00:39:44

07/07/2021 05:04 PM
ACCEPTED: X
REJECTED:
Operator: U01