
Midmark M11 - v1.0.0
Total Cycles: 2010

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2022 01:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 180.8 F 4.4
4:00 189.0 F 8.8
6:00 207.9 F 9.2
8:00 224.2 F 8.4
10:00 237.1 F 12.5
12:00 238.0 F 10.7
14:00 236.1 F 8.5
16:00 249.8 F 15.4
18:00 261.6 F 22.1
20:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.3 F 29.0
2:00 271.5 F 28.7
3:00 272.3 F 29.1
4:00 271.3 F 28.5

Min 271.3 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 222.2 F 0.0
10:00 234.7 F 0.0
15:00 236.1 F 0.0
20:00 229.5 F 0.0
25:00 223.8 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

06/08/2022 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____