
Midmark M11 - v1.0.0
Total Cycles: 2065

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2022 06:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.5 F 0.0
2:00 85.0 F 0.0
4:00 87.4 F 0.0
6:00 136.7 F 0.7
8:00 165.7 F 2.6
10:00 191.0 F 5.0
12:00 204.6 F 8.1
14:00 225.8 F 11.4
16:00 225.5 F 5.3
18:00 233.1 F 9.3
20:00 243.0 F 13.4
22:00 237.4 F 10.0
24:00 236.4 F 8.3
26:00 246.4 F 13.3
28:00 256.9 F 19.2
30:00 267.0 F 25.3
32:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 271.9 F 29.4
2:00 272.5 F 29.4
3:00 272.9 F 29.2
4:00 272.0 F 29.0

Min 271.3 F 29.0
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.8 F 0.0
5:00 230.8 F 0.0
10:00 185.5 F 0.0
15:00 195.2 F 0.0
20:00 195.7 F 0.0
25:00 232.7 F 0.0
30:00 224.0 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 32:22
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:10:20
CYCLE COMPLETE

06/28/2022 08:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____