
Midmark M11 - v1.0.0
Total Cycles: 2629

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2022 01:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.3 F 0.0
2:00 181.9 F 4.6
4:00 197.7 F 9.3
6:00 216.9 F 7.9
8:00 221.2 F 8.0
10:00 234.1 F 13.2
12:00 237.1 F 10.3
14:00 235.0 F 8.2
16:00 249.1 F 15.1
18:00 261.1 F 21.9
20:00 272.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.5
2:00 271.7 F 28.8
3:00 271.6 F 29.2
4:00 272.3 F 29.4

Min 271.2 F 28.6
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 221.0 F 0.0
10:00 232.7 F 0.0
15:00 230.1 F 0.0
20:00 229.9 F 0.0
25:00 229.8 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

11/30/2022 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____