

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000089  
10:22 28/08/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:59  
PRESSURIZING 3:02  
STERILIZING 4:26  
136.5 C 225kPa 4:26  
136.4 C 225kPa 4:56  
136.3 C 225kPa 5:26  
136.4 C 225kPa 5:56  
136.4 C 225kPa 6:26  
136.4 C 226kPa 6:56  
136.4 C 224kPa 7:26  
136.2 C 224kPa 7:56  
136.2 C 224kPa 7:56  
VENTING 7:57  
AIR DRYING 8:51  
CYCLE COMPLETE 23:51  
-----

Digital Signature #  
00FEC6047256B792  
-----